

Certificate of Analysis

eShop Dummy Customer US
used in eShop only
BLANCHARD
USA

Print Date: 23-Feb-2021

Product Name: MgCl₂, 10mM Solution For
Use with LAL, 30 ml vial
Material Number: S50-641
Batch No: 0000904575
Manufacturing Date: 17-Jun-2020
Expiration Date: 17-Jun-2023

<i>Test</i>	<i>RESULT</i>	<i>SPECIFICATION</i>		<i>UNIT</i>
		<i>MIN</i>	<i>MAX</i>	
Sterility Test	Negative			
Endotoxin Content (EU/ml)	< 0.005	***	<0.005	EU/ml
1 EU/ml = 1 IU/ml				
Wet ID: Passes wet test for magnesium	Pass			
Wet ID: Passes wet test for chloride	Pass			

Additional Information:

This product was tested on a non tissue culture treated plate.

Raj Patel

Electronically signed by Raj Patel

Date: 29-JUL-2020 16:14:33 EST

RELEASE (Inspection Lot: Usage Decision)

This lot has been reviewed by Quality Assurance in compliance with requirements of Lonza's Quality System. This document was generated from a validated Part 11-compliant electronic system and thus handwritten signatures are not required.

For Technical Assistance, call 1-800-521-0390

Lonza Walkersville Inc.
8830 Biggs Ford Road
Walkersville, MD 21793 8415
Tel (301) 898 7025
Fax (301) 845 4024