

**LAL Division**

1600 Bellwood Road
Phone: (804) 271-7677,
Fax: (804) 271-7791
Richmond, VA 23237 USA

Certificate of Analysis**Lysate Reagent Water (LRW)**

| | |
|------------------------------------|--|
| Product Code: | LRW-2030 |
| Product Description: | Lysate Reagent Water (LRW) 30 mL/vial |
| Lot Number: | 031803 |
| Production Date: | 15 MAR 2018 |
| Quality Assurance Release Date: | 21 MAY 2019 |
| Expiration Date: | 2022.03 |
| <u>Assay:</u> | <u>Result:</u> |
| Bacterial Endotoxin Test | <0.001 EU/mL |
| Interfering Test | Pass |

This is to certify that the above lot meets Wako Chemicals USA, Inc. – LAL Division's product specifications and complies with all applicable requirements for product release. (Less than 0.001 EU/mL)

Created By: *Jason Walden* Date: 05/21/2019

Quality Assurance
Approved By: *Jason Holliday* Date: 05/21/2019