

LAL Division
1600 Bellwood Road
Richmond, VA 23237 USA
Phone: (804) 271-7677,
Fax: (804) 271-7791

Page 1 of 1

Certificate of Analysis

LPS for Endotoxin Indicator, >100K EU/vial

Product Code:	WLPS-0100K
Product Description:	LPS for Endotoxin Indicator
Lot Number:	032012
Production Date:	29 MAR 2020
Expiration Date:	2021.03

<u>Assay:</u>	<u>Specification:</u>	<u>Result:</u>
Endotoxin Potency (EU per Vial)	>100K EU/vial	Pass

Reference

United States Pharmacopeia (USP) 2019, General Chapters: <1228> Depyrogenation, <1228.1> Dry heat depyrogenation, and <1228.5> Endotoxin indicators for depyrogenation.

This is to certify that this product lot was manufactured according to current cGMP Standards, meets FUJIFILM Wako Chemicals U.S.A. Corporation – LAL Division's product specifications and complies with all applicable requirements for product release.

Created By: Jason HollidayDate: 4/7/2020Quality Assurance Approval By: Jack WebbDate: 4/7/2020