
CERTIFICATE OF ANALYSIS

Product Code: W50-100
Product: LAL Reagent Water - 100ml

Lot Number: 0000713422
Manufacture Date: 18-May-2018
Expiration Date: 04-May-2020

TEST (Method)

SPECIFICATIONS

	Min.	Max.	Results
Sterility Test	Negative	***	Negative
Endotoxin-Media (EU/ml)	<0.005	***	< 0,005
PPC Recovery (%)	75	150	78 %

This lot has been reviewed by Quality Assurance in compliance with requirements of Lonza's Quality System.

This document was generated from a validated Part 11-compliant electronic system and thus handwritten signatures are not required.