
CERTIFICATE OF ANALYSIS

Product Code: W50-640
Product: LAL Reagent Water - 30ml

Lot Number: 0000742504
Manufacture Date: 19-Sep-2018
Expiration Date: 19-Mar-2021

TEST (Method)

SPECIFICATIONS

	Min.	Max.	Results
Sterility Test	Negative	***	Negative
Endotoxin Content (EU/ml)	***	< 0.005	< 0,005
1 EU/ml = 1 IU/ml			
PPC Recovery (%)	75	150	110 %

This lot has been reviewed by Quality Assurance in compliance with requirements of Lonza's Quality System.

This document was generated from a validated Part 11-compliant electronic system and thus handwritten signatures are not required.