Lonza Walkersville Inc. 8830 Biggs Ford Road Walkersville, MD 21793 8415 Tel (301) 898 7025 Fax (301) 845 4024



Printed on, 28-May-2019 15:00

Page 1/ 1

## **CERTIFICATE OF ANALYSIS**

**Product Code:** 

W50-100

Product:

LAL Reagent Water - 100ml

Lot Number:

0000696509

Manufacture Date:

28-Feb-2018

**Expiration Date:** 28-Feb-2020

TEST (Method)	SPECIFICATIONS			
	Min.	Max.	Results	
Sterility Test	Negative	***	Negative	
Endotoxin-Media (EU/ml)	<0.005	***	< 0,005	
PPC Recovery (%)	75	150	129 %	

This lot has been reviewed by Quality Assurance in compliance with requirements of Lonza's Quality System.

This document was generated from a validated Part 11-compliant electronic system and thus handwritten signatures are not required.